

**Specification :**

- Consist of three concentric pipes
- Thermocouples at proper position are used to measure the temperature.
- Inside the inner pipe a nicrome wire heater is placed.
- Heat input to the heater is given through a variac and measured by a voltmeter and ammeter.

Note: Specifications are subject to change, Photos shown above are Indicative, Actual Product can Vary.



Export Sales: +91-9829132777
India Sales: +91-9588842361



IT-2013, Ramchandrapura Industrial Area,
Sitapura Extension, Jaipur-302022, India.



info@tesca.in
www.tescaglobal.com