

**Specifications:**

- Variable Speed Fan.
- Outside Diameter Of The Test Tube Between 30 Mm And 40 Mm.
- Tube Thickness Between 1mm And 1.5mm.
- External Diameter Of The External Tube Of The Exchanger Between 60 Mm And 80 Mm.
- Test Tube Length Between 1400 Mm And 1700 Mm.
- Variable Heating Power.
- Heater Power Reader.
- Thermocouples And Temperature Indicator Along The Test Tube (minimum 12 Indicators).
- The Test Bench Must Be Equipped With Pressure Gauges And Diaphragms.
- Test Bench Must Be Mounted On A Frame With Wheels. User Manual And Practical Work

Note: Specifications are subject to change, Photos shown above are Indicative, Actual Product can Vary.



Export Sales: +91-9829132777
India Sales: +91-9588842361



IT-2013, Ramchandrapura Industrial Area,
Sitapura Extension, Jaipur-302022, India.



info@tesca.in
www.tescaglobal.com